

IFLOWROCKER

User Guide



Simple Platform for Complex Biology

Congratulations on your purchase of an IFlowRocker™ by OrganoBiotech. Please read this user guide which explains proper operation. This manual is also posted on our website, www.organo-biotech.com. We're confident that your IFlowRocker will become an essential tool in your laboratory and we wish you success with your work.



1. SYMBOLS USED ON THE IFLOWROCKER



Caution: Follow the Instructions in the Operator's Manual



Indicates European Conformity (Conformité Européene) with health, safety, and environmental protection standards



Please dispose of the IFlowRocker in accordance with local regulations

2. SCREEN ICONS



Home



Maintenance



Start / Resume



Information



Edit



Stop



Settings



Save



Pause

3. OVERVIEW

The IFlowRocker is a programmable tilt table. You can program it to travel at a specified rate to a specified angle, and at a second specified rate to a second angle, and optionally to one or two more angular positions at specified rates, and even have the platform pause at those tilt angles for seconds or even hours. This lab instrument is designed for very accurate angular position and speed control. You can even program it to travel very slowly to offset the changing liquid levels in the inlet and outlet reservoirs of microfluidic systems, thereby providing a constant flow rate of medium.

Operational space required (+\ -30 degrees tilt and single shelf):
minimal height of 20 cm; 37.5cm of width; and 39cm in depth
+\ -30 degrees

4. IQ/OQ/PQ

- The Installation Qualification (IQ) consists of plugging in the rocker, turning on the power, and checking if the display lights up.
- The Operation Qualification (OQ) consists of going to the Maintenance screen (see below) and running the test program.
- The Performance Qualification (PQ) consists of running the IFlowRocker for two cycles of a program which includes the slowest pace, the fastest pace, and the greatest angle, at which you will use the IFlowRocker.

5. SETUP

Place the IFlowRocker on a stable, level lab bench or shelf in an incubator. Carry it by grasping the bottom sides. Plug the AC/DC power adapter into the 24V socket on the back of the IFlowRocker and then insert the plug into a wall outlet. Turn on the power and the display will light up, initially with a splash screen, then with the home screen. It is now ready to be used

6. INFORMATION

The Information screen displays the model and software revision

7. SETTING

Click on the Settings icon to jump to the Settings screen and program the IFlowRocker. On this screen you can select any of 6 programs and see their settings. Once a program is selected you can edit the program by clicking on the Edit icon.

To edit a program, use the keypad on the right. Note that the rate of travel is actually programmed as pace in seconds per degree. This is done to make it easy to determine how long it will take to travel to a specified angle. Programming travel rate as pace also makes it very easy to program slow rates of the change in tilt angle to provide for constant flow of culture medium in microfluidics applications. Once the values are entered, click on Save. Unless you click on Save or Revert, you will not be able to leave this screen.

8. MAINTENANCE

On the Maintenance screen you can set the level position and run the test program. To set the level position, use the arrows to tilt the platform to the desired position. Note that as you press an arrow, the rate of travel increases. Use the double arrow to travel more quickly, the single arrow to travel more slowly. Once the platform is at the desired angular position, click on the Set Level button.

Click on the Test Program button to run the test program. You can use this to check if the rocker is working including for Operation Qualification. This program is set in the factory and cannot be edited.

9. OPERATION

To begin using your IFlowRocker, select the program you want to use. If you want to change the program's settings, go to the SETTINGS screen, change the settings, save the program, and

return to the home screen. The next step is to have the rocker advance to the initial tilt position. Once the platform is in the initial tilt position, press START. Once the rocker is running, you can PAUSE, and then RESUME or STOP.

10. CLEANING

If you wish to clean your IFlowRocker, clean the outside of the unit only with mild soap, water and a soft cloth. Under normal conditions, the IFlowRocker should never need to be disassembled for cleaning.

11. SUPPORT

FAQs and other helpful information are available on our website, www.organo-biotech.com. If you cannot find an answer there, please email info@organo-biotech.com.

12. SPECIFICATIONS

Power Requirement: 24 VDC, 2.5 Amp

Relative Humidity: 5 – 90% non-condensing

Operating Temperature: 4 - 40°C

Altitude: <2000m

Storage Temperature: -40 to 50°C

Meets **CE** requirements

13. WARRANTY

OrganoBiotech warrants its Products against defects in materials and workmanship for time periods which vary according to the Product. Within these time periods, OrganoBiotech will replace or repair, without charge to the original purchaser, any part which is defective. The IFlowRocker warranty is one year. The warranty is void if the Product is defective due to product accident, product modification, exposure to radiation other than for sterilization, connection to an improper electrical supply, lack of proper maintenance, contamination, improper installation or misuse. If the product is used in a manner not specified by the manufacturer, the

protection provided by the equipment may be impaired. The warranty shall also not apply to defects arising from fire, flood, lightning or other conditions unrelated to correct operation of the Product.

OrganoBiotech's liability is limited, at the company's election, to (1) refund of the original purchaser's purchase price for the Product (2) repair of the Product, or (3) replacement of the Product or defective parts. Evidence of purchase by the original purchaser is required. OrganoBiotech may also request documentation of proper maintenance, if applicable.

Operator is responsible for: providing proof of purchase and providing normal care and maintenance.

14. WARNINGS AND CAUTIONS

- Read the user's manual before operating.
- Do not put excessive weight on the platform.
- Do not insert fingers or objects under the platform.
- When working with hazardous or pathogenic samples, operate the IFlowRocker in a biosafety cabinet or other standard laboratory safety enclosure.
- No user serviceable parts are inside of the instrument.
- For indoor use only.
- Pollution Degree 2 per EN 61010-1.
- Overvoltage Category II per EN 61010-1.
- Enclosure Protection: Not Protected Against the Ingress of Moisture.
- Do not immerse in liquid.
- Before touching the IFlowRocker, touch a bare metal surface to discharge static electricity.

15. DISCLAIMER

OrganoBiotech makes no other warranty, expressed or implied, with respect to its Products. ORGANOBIOTECH MAKES NO WARRANTY RESPECTING THE MERCHANTABILITY OF THE PRODUCTS OR THEIR SUITABILITY OR FITNESS FOR ANY PARTICULAR PURPOSE OR USE. In no event shall OrganoBiotech be liable for indirect, special, incidental or consequential damages of any nature. OrganoBiotech Inc is not liable for any damages, including but not limited to, lost profits,

lost savings, or other incidental or consequential damages arising from ownership or use of this product, or for any delay in the performance of its obligations under the warranty due to causes beyond its control. Any recovery for any claim shall be limited to the original purchase price for the product.

OrganoBiotech also reserves the right to make any improvements or modifications to the product described in this manual at any time, without notice of these changes.

OrganoBiotech products are not designed, intended, or authorized for use in applications or as system components intended to support or sustain human life, as a clinical medical device for humans, or for any application in which the failure of the product could create a situation where personal injury or death may occur.

16. CONTACT INFO

OrganoBiotech Inc, Oakville, ON, CANADA

www.organo-biotech.com

info@organo-biotech.com



Simple Platform for Complex Biology